



UK Declaration of Conformity

For the following equipment :

Product Name: Switching Power Supply

Model Designation: RPX-75y (x=S,D,T) (y=-3.3,-5,-12,-15,-24,-36,-48,03,A,B,C,D,03)

The designated product(s) is(are) in conformity with the relevant legislation:

The Restriction of the Use of Certain Hazardous Substances in Electrical and Electronic Equipment Regulations 2012: SI 2012 No. 3032

Medical Devices Regulations 2002 (SI 2002 No 618, as amended) (UK MDR 2002)

BS EN 60601-1:2006+ A12:2014

TUV certificate No : TA50220730

Medical Devices Regulations 2002 (SI 2002 No 618, as amended) (UK MDR 2002)

BS EN60601-1-2:2015

EMI (Electro-Magnetic Interference)

Conducted emission / Radiated emission

BS EN 55011:2016+A1:2017

Class B

Harmonic current

BS EN 61000-3-2:2014

Voltage flicker

BS EN 61000-3-3:2013

EMS (Electro-Magnetic Susceptibility)

ESD air

BS EN 61000-4-2:2009

Level 4

15KV

ESD contact

BS EN 61000-4-2:2009

Level 4

8KV

RF field susceptibility

BS EN 61000-4-3:
2006+A1:2008+A2:2010

Level 3

10V/m(80MHz~2.7GHz)

EFT bursts

BS EN 61000-4-4:2012

Level 3

9~28V/m(385MHz~5.78GHz)

Surge susceptibility

BS EN 61000-4-5:2014 +A1:2017

Level 4

2KV/100KHz

Surge susceptibility

BS EN 61000-4-5:2014 +A1:2017

Level 4

4KV/Line-Earth

Conducted susceptibility

BS EN 61000-4-6:2014

Level 3

10V

Magnetic field immunity

BS EN 61000-4-8:2010

Level 4

30A/m

Voltage dip, interruption

BS EN 61000-4-11:2004+A1:2017

>95% dip 0.5 periods

30% dip 25 periods

>95% interruptions 250 periods

Note:

The power supply is considered as a component that will be operated in combination with final equipment. Since EMC performance will be affected by the complete system, the final equipment manufacturers must re-qualify EMC Regulations on the complete system again.

The EMC tests mentioned above are performed using a well defined metal plate to simulate said metal enclosure.

For guidance on how to perform these EMC tests, please refer to "EMI testing of component power supplies".(as available on <http://www.meanwell.com>)".

The product under declaration is just a unit without medical function. Complete MDR should only be verified when it is used together with particular medical device(s).

This Declaration is effective from serial number SC1xxxxxxx

Person responsible for marking this declaration :

MEAN WELL Enterprises Co., Ltd.

(Manufacturer Name)

No.28, Wuquan 3rd Rd., Wugu Dist., New Taipei City 24891, Taiwan

(Manufacturer Address)

Aries Jian/ Director, Group R&D :

(Name / Position)

(Signature)

Taiwan

(Place)

Alex Tsai/ Director, Product Strategy Center :

(Name / Position)

(Signature)

Jun. 28th, 2021

(Date)