



Ref. Certif. No.

**DK-161649-UL**

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

**CB TEST CERTIFICATE**

Product

AC/DC Medical Adapter

Name and address of the applicant

MEAN WELL ENTERPRISES CO., LTD.  
No 28 Wuquan 3rd Rd  
Wugu District New Taipei City 24891  
Taiwan

Name and address of the manufacturer

MEAN WELL ENTERPRISES CO., LTD.  
No 28 Wuquan 3rd Rd  
Wugu District New Taipei City 24891  
Taiwan

Name and address of the factory

MEAN WELL Enterprises Co., Ltd.  
No.28, Wuquan 3rd Rd., Wugu District, New Taipei City 24891,  
Taiwan

Note: When more than one factory, please report on page 2

☒ Additional Information on page 2

Ratings and principal characteristics

Input: 100-240 Vac, 50/60 Hz,  
1.0-0.5 A for GSM40Ax series,  
1.4-0.7 A for GSM60Ax series  
Output: See test report for details.

Trademark / Brand (if any)



Mean Well

Customer's Testing Facility (CTF) Stage used

CTF Stage 2

Model / Type Ref.

GSM60Ax, GSM40Ax  
☒ Additional Information on page 2

Additional information (if necessary may also be reported on page 2)

The risk management requirements of the standard were not addressed  
National Differences: CA, EU Group Differences, JP, US  
☒ Additional Information on page 2

A sample of the product was tested and found to be in conformity with

IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012, IEC 60601-1:2005/AMD2:2020

As shown in the Test Report Ref. No. which forms part of this Certificate

E227340-D6002-CB-1 issued on 2024-12-27

This CB Test Certificate is issued by the National Certification Body



- ☐ UL Solutions (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- ☒ UL Solutions (Denko), Borupvang 5A DK-2750 Ballerup, DENMARK
- ☐ UL Solutions (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- ☐ UL Solutions (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see [www.ul.com/ncbnames](http://www.ul.com/ncbnames)

Date: 2024-12-27

Signature:

Thomas Wilson



Ref. Certif. No.

**DK-161649-UL**

**Factory(ies):**

SuZhou MEAN WELL Technology Co., Ltd.  
No.269, Changping Rd, Huangdai Town Xiangcheng District, Suzhou, Jiangsu, 215152,  
China

MEAN WELL (GUANGZHOU) ELECTRONICS CO., LTD  
No.11 Jingu South Road,Huadu District, Guangzhou Guangdong 510890,  
China

MEAN WELL INDIA ELECTRONICS PRIVATE LIMITED  
9C, Peenya industrial area, Chokkasandra, 2ND PHASE,  
PEENYA, Bengaluru (Bangalore) Urban, Karnataka, 560058,  
India

YONGDEN TECHNOLOGY CORP  
345 MacArthur Hwy, Tabang, Guiguinto, Bulacan 3015,  
Philippines

**Additional Model Detail(s):**

GSM40Ax, (x can be 05, 07, 09, 12, 13.5-KOD, 15, 18, 24, 28 or 48)  
GSM60Ax, (x can be 05, 07, 09, 12, 15, 18, 24 or 48)

**Additionally evaluated to:**

EN 60601-1:2006, EN 60601-1:2006/A1:2013, EN 60601-1:2006/A12:2014, EN 60601-1:2006/A2:2021

**Additional information (if necessary)**



- ☐ UL Solutions (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- ☒ UL Solutions (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- ☐ UL Solutions (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- ☐ UL Solutions (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see [www.ul.com/ncbnames](http://www.ul.com/ncbnames)

Date: 2024-12-27

Signature:

Thomas Wilson