

DE 2-040830

## IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

## **CB TEST CERTIFICATE**


Product	Medical Type Switching Power Supply
Name and address of the applicant	MEAN WELL Enterprises Co., Ltd. No. 28, Wuquan 3rd Rd., Wugu District, New Taipei City 24891 Taiwan
Name and address of the manufacturer	MEAN WELL Enterprises Co., Ltd. No. 28, Wuquan 3rd Rd., Wugu District, New Taipei City 24891 Taiwan
Name and address of the factory	See additional page(s)
Ratings and principal characteristics	AC 100-240V; 50/60Hz; 1.3-0.6A; Class I; for output ratings refer to test report page 9
Trademark (if any)	Trademark of MEAN WELL
Customer's Testing Facility (CTF) Stage used	N/A
Model / Type Ref.	GSM90Ax (x= 12, 15, 19, 24 or 48 to denote different output ratings)
Additional information (if necessary may also be reported on page 2)	Models differ in model name, output rating, windings of main transformer T1 and the ratings of components; The risk management and usability requirements of the standard were not addressed; for details see test report
A sample of the product was tested and found to be in conformity with	IEC 60601-1:2005+A1+A2 for national differences see test report
As shown in the Test Report Ref. No. which forms part of this Certificate	CN2387YB 001
This CP Test Contificate is issued by the National Contification Backs	

This CB Test Certificate is issued by the National Certification Body



Disclaimer: This is an electronically released document. The authenticity of this certificate can be verified on the IECEE Website "http://certificates.iecee.org"

DE 2-040830



10/061a CB 06/20v3 rl

Disclaimer: This is an electronically released document. The authenticity of this certificate can be verified on the IECEE Website "http://certificates.iecee.org"