

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT
(IECEE) CB SCHEME

CB TEST CERTIFICATE

Product	AC/DC Switching Adaptor
Name and address of the applicant	MEAN WELL Enterprises Co., Ltd. No.28, Wuquan 3rd Rd, Wugu District, 24891 New Taipei City, Taiwan
Name and address of the manufacturer	MEAN WELL Enterprises Co., Ltd. No.28, Wuquan 3rd Rd, Wugu District, 24891 New Taipei City, Taiwan
Name and address of the factory	SuZhou MEAN WELL Technology Co., Ltd. No.269, Changping Road, Huangdai Town, Xiangcheng District, 215152 Suzhou Jiangsu, China
<i>When more than one factory</i>	<input checked="" type="checkbox"/> Additional information on page 2
Ratings and principal characteristics	Product data Ratings: Input : 100-240Vac, 50/60Hz, 0,5-0,25A Output: see test report for details
Trademark / Brand (if any)	 MEAN WELL
Customer's Testing Facility (CTF) Stage used	
Model / Type Ref.	NGE12xyzzzz and NGE18xyzzzz
Additional information	<input checked="" type="checkbox"/> Additional information on page 2
A sample of the product was tested and found to be in conformity with	IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012, IEC 60601-1:2005/AMD2:2020, IEC 60601-1-11:2015 and IEC 60601-1-11:2015/AMD1:2020 National differences: AU, EU Group Differences, EU Special National Conditions, GB, JP, NZ, SA Comments: This CB test certificate replaces CB test certificate NL-103326.
As shown in the Test Report Ref. No. which forms part of this Certificate	4915470.50, 4915470.51

This CB Test Certificate is issued by the National Certification Body

DEKRA Certification B.V.
Meander 1051
6825 MJ Arnhem
Netherlands

Date: 7 April 2024

Signature: R Zhou

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT
(IECEE) CB SCHEME

Additional factory

MEAN WELL (GUANGZHOU) Electronics Co., Ltd.
No.11 Jingu South Road, Huadu District, 510890 Guangzhou Guangdong, China

MEAN WELL Enterprises Co., Ltd.
No.28, Wuquan 3rd Rd, Wugu District, 24891 New Taipei City, Taiwan

YONGDEN TECHNOLOGY CORPORATION
345 MacArthur HighWay Tabang, 3015 Guiguinto, Bulacan, Philippines

MEAN WELL INDIA ELECTRONICS PRIVATE LIMITED
9C, Peenya industrial area, Chokkasandra, 2ND PHASE, PEENYA, 560058 Bengaluru (BANGALORE) Urban,
Karnataka India

Additional information

Only 4915470.50 considers JP deviations.

Model reference:

x=E, U, I, UK, AU, CN, KR, IN, y=05, 09, 12, 15, 18, 24 zzzz= maybe Blank, -,0~9,A~Z or a~z for market purpose
except models NGE 12x05-USB

The risk management requirements were not addressed.
The Usability evaluation has not been addressed.

This CB Test Certificate is issued by the National Certification Body

DEKRA Certification B.V.
Meander 1051
6825 MJ Arnhem
Netherlands



Date: 7 April 2024

Signature: R Zhou

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT
(IECEE) CB SCHEME**CB TEST CERTIFICATE**

Product	AC/DC Switching Adaptor
Name and address of the applicant	MEAN WELL Enterprises Co., Ltd. No.28, Wuquan 3rd Rd, Wugu District, 24891 New Taipei City, Taiwan
Name and address of the manufacturer	MEAN WELL Enterprises Co., Ltd. No.28, Wuquan 3rd Rd, Wugu District, 24891 New Taipei City, Taiwan
Name and address of the factory	SuZhou MEAN WELL Technology Co., Ltd. No.269, Changping Road, Huangdai Town, Xiangcheng District, 215152 Suzhou Jiangsu, China
<i>When more than one factory</i>	<input checked="" type="checkbox"/> Additional information on page 2
Ratings and principal characteristics	Product data Ratings: Input : 100-240Vac, 50/60Hz, 0,5-0,25A Output: see test report for details
Trademark / Brand (if any)	 MEAN WELL
Customer's Testing Facility (CTF) Stage used	
Model / Type Ref.	NGE 12x05-USB
Additional information	<input checked="" type="checkbox"/> Additional information on page 2
A sample of the product was tested and found to be in conformity with	IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 National differences: AU, EU Group Differences, EU Special National Conditions, GB, JP, NZ, SA Comments: This CB test certificate replaces CB test certificate NL-103322.
As shown in the Test Report Ref. No. which forms part of this Certificate	4915470.50

This CB Test Certificate is issued by the National Certification Body

DEKRA Certification B.V.
Meander 1051
6825 MJ Arnhem
Netherlands

Date: 7 April 2024

Signature: R Zhou

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT
(IECEE) CB SCHEME

Additional factory

MEAN WELL (GUANGZHOU) Electronics Co., Ltd.
No.11 Jingu South Road, Huadu District, 510890 Guangzhou Guangdong, China

MEAN WELL Enterprises Co., Ltd.
No.28, Wuquan 3rd Rd, Wugu District, 24891 New Taipei City, Taiwan

YONGDEN TECHNOLOGY CORPORATION
345 MacArthur HighWay Tabang, 3015 Guiguinto, Bulacan, Philippines

MEAN WELL INDIA ELECTRONICS PRIVATE LIMITED
9C, Peenya industrial area, Chokkasandra, 2ND PHASE, PEENYA, 560058 Bengaluru (BANGALORE) Urban,
Karnataka India

Additional information

Model reference:
x=E, U, I, UK, AU, CN, KR, IN

Only the requirements of IEC 60601-1, as documented in the accompanying test report, have been addressed.
The risk management requirements were not addressed.
The Usability evaluation has not been addressed.

This CB Test Certificate is issued by the National Certification Body

DEKRA Certification B.V.
Meander 1051
6825 MJ Arnhem
Netherlands



Date: 7 April 2024

Signature: R Zhou