CERTIFICATE

Issued to: Applicant: MEAN WELL Enterprises Co., Ltd. No.28, Wuquan 3rd Rd, Wugu District 24891 New Taipei City, Taiwan

Licensee:

MEAN WELL Enterprises Co., Ltd. No.28, Wuquan 3rd Rd, Wugu District 24891 New Taipei City, Taiwan

Product	:	AC/DC Switching Adaptor
Trade name(s)	:	MEAN WELL
Type(s)/model(s)	:	NGE30xyzzzz

The product and any acceptable variation thereto as specified in the Annex to this certificate and the documents therein referred to.

DEKRA hereby declares that the above-mentioned product has been certified on the basis of

- a type test according to EN 60601-1:2006, EN 60601-1:2006/A1:2013, EN 60601-1:2006/A2:2021, EN 60601-1-11:2015 and EN 60601-1-11:2015/A1:2021
- an inspection of the factory location according to CENELEC Operational Document CIG 021
- a DEKRA certification agreement with the number 6059480

DEKRA hereby grants the right to use the DEKRA Mark.

The DEKRA Mark may be applied to the product as specified in this certificate for the duration and under the conditions of the DEKRA Mark certification agreement.

This certificate is issued on 25 January 2024 and expires upon withdrawal of one of the above mentioned standards.

Certificate number: 35-134873

DEKRA Certification B.V.

B.T.M. Holtus Managing Director

C. Lin Certification Manager

© Integral publication of this certificate is allowed

ACCREDITED BY THE DUTCH ACCREDITATION COUNCIL



DEKRA Certification B.V. Meander 1051, 6825 MJ Arnhem P.O. Box 5185, 6802 ED Arnhem, Netherlands T +31 88 96 83000 F +31 88 96 83100 Company registration 09085396

DEKRA



ANNEX TO DEKRA Mark CERTIFICATE 35-134873

page 1 of 2

SPECIFICATION OF THE CERTIFIED PRODUCT

Product data

Product Trade name(s) Type(s)/model(s) Rated input voltage Rated frequency Rated input current AC/DC Switching Adaptor
MEAN WELL
NGE30xyzzzz
100-240Vac
50/60Hz
1,0-0,6A

TESTS

Test requirements

EN 60601-1:2006 EN 60601-1:2006/A1:2013 EN 60601-1:2006/A2:2021 EN 60601-1-11:2015 EN 60601-1-11:2015/A1:2021

Test result

The test results are laid down in DEKRA test file 490815000.

Additional information

Model reference: x=E,I, y=05, 09, 12, 15, 18, 24, 48, zzzz=maybe Blank, -,0~9,A~Z or a~z for market purpose

The list of components is laid down in test report 4908150.53, 4908150.54.

Conclusion

The examination proved that all requirements were met.

Factory location

SuZhou MEAN WELL Technology Co., Ltd. No.269, Changping Road, Huangdai Town, Xiangcheng District 215152 Suzhou Jiangsu, China



ANNEX TO DEKRA Mark CERTIFICATE 35-134873

page 2 of 2

Trade name(s) :

