



## UK Declaration of Conformity

For the following equipment :

Product Name: Medical Type Switching Power Supply

Model Designation: RPSY-160Z (Y=G or blank) (Z=-5,-12,-15,-24,-48)

The designated product(s) is(are) in conformity with the relevant legislation:

### The Restriction of the Use of Certain Hazardous Substances in Electrical and Electronic Equipment Regulations 2012: SI 2012 No. 3032

### Medical Devices Regulations 2002 (SI 2002 No 618, as amended) (UK MDR 2002)

BS EN 60601-1: 2006+A12:2014

TUV certificate No : TA50147896

### Medical Devices Regulations 2002 (SI 2002 No 618, as amended) (UK MDR 2002)

BS EN 60601-1-2: 2015

#### EMI (Electro-Magnetic Interference)

Conducted emission / Radiated emission

BS EN 55011:2016+A1:2017

Class B

Harmonic current BS EN 61000-3-2:2014

Voltage flicker BS EN 61000-3-3:2013

#### EMS (Electro-Magnetic Susceptibility)

ESD air BS EN 61000-4-2:2009 Level 4 15KV

ESD contact BS EN 61000-4-2:2009 Level 4 8KV

RF field susceptibility BS EN 61000-4-3: 2006+A1:2008+A2:2010 Level 3 10V/m(80MHz-2.7GHz)

RF field susceptibility BS EN 61000-4-3: 2006+A1:2008+A2:2010 Table 9 9~28V/m(385MHz~5.78GHz)

EFT bursts BS EN 61000-4-4:2012 Level 3 2KV/100KHz

Surge susceptibility BS EN 61000-4-5:2014+A1:2017 Level 4 2KV/Line-Line

Surge susceptibility BS EN 61000-4-5:2014+A1:2017 Level 4 4KV/Line-Earth

Conducted susceptibility BS EN 61000-4-6:2014 Level 3 10V

Magnetic field immunity BS EN 61000-4-8:2010 Level 4 30A/m

Voltage dip, interruption BS EN 61000-4-11:2004 +A1:2017 >95% dip 0.5 periods 30% dip 25 periods >95% interruptions 250 periods

#### Note:

The power supply is considered as a component that will be operated in combination with final equipment. Since EMC performance will be affected by the complete system, the final equipment manufacturers must re-qualify EMC Regulations on the complete system again.

The EMC tests mentioned above are performed using a well defined metal plate to simulate said metal enclosure.

For guidance on how to perform these EMC tests, please refer to "EMI testing of component power supplies".(as available on <http://www.meanwell.com>).

The product under declaration is just a unit without medical function. Complete MDR should only be verified when it is used together with particular medical device(s).

This Declaration is effective from serial number SC1xxxxxxx

#### Person responsible for marking this declaration :

MEAN WELL Enterprises Co., Ltd.

(Manufacturer Name)

No.28, Wuquan 3rd Rd., Wugu Dist., New Taipei City 24891, Taiwan

(Manufacturer Address)

Aries Jian/ Director, Group R&D :

(Name / Position)

(Signature)

Alex Tsai/ Director, Product Strategy Center :

(Name / Position)

(Signature)

Taiwan

(Place)

Jun. 28th, 2021

(Date)