

# Technical Compliance Statement

## EMC Test Report

### For the following information


Ref. File No.: C1M1604001

Product : Switching Power Supply  
Model Number : RPS-300-x-y (x=12, 15, 24, 27, 48; y=C or Blank)  
Applicant : MEAN WELL Enterprises Co., Ltd.  
Manufacturer #1 : MEAN WELL Enterprises Co., Ltd.  
Manufacturer #2 : MEAN WELL (GUANGZHOU) Electronics Co., Ltd.  
Manufacturer #3 : SuZhou MEAN WELL Technology Co., Ltd.  
Standards :  
EN 60601-1-2:2015  
EN 55011:2009 +A1:2010  
EN 61000-3-2:2014, EN 61000-3-3:2013  
EN 61000-4-2:2009, EN 61000-4-3:2006 +A1:2008 +A2:2010, EN 61000-4-4:2012,  
EN 61000-4-5:2014, EN 61000-4-6:2014, EN 61000-4-8:2010,  
EN 61000-4-11:2004

We hereby certify that the above product has been tested by us with the listed standards and found in compliance with the council EMC directive 93/42/EEC.

The test data & results are issued on the EMC test report no. **EM-E160247**.

Signature

  
\_\_\_\_\_  
Alex Deng/Deputy Manager  
Date: 2016. 06. 07

Test Laboratory:  
AUDIX Technology Corporation, EMC Department  
TAF Accreditation No.: 1724  
Web Site: [www.audixtech.com](http://www.audixtech.com)



The statement is based on a single evaluation of one sample of the above-mentioned products. It does not imply an assessment of the whole production and does not permit the use of the test lab logo.